E2M0.7 & E2M1.5 vacuum pumps

Regular maintenance of your Edwards EM series vacuum pump will ensure safe operation and optimised performance. Edwards offers service kits, oils and component spares for complete pump/ accessory overhaul and blade replacement.

Please refer to the maintenance schedule for recommended maintenance guidelines and service intervals. Corrosive or abrasive applications, or the use of gas ballast, may increase service intervals.



Performance optimisation through regular maintenance

Oil condition

The condition of oil used in mechanical pumps deteriorates over time and by contamination from process gases and byproducts, indicated by a progressive change of oil colour. Regular replacement of oil will minimise wear and tear on contacting parts and prolong good vacuum performance.

Edwards recommends regular checks on the oil colouration and replacement every 3,000 hours (4 months) depending on application.

Clean and overhaul

Regular cleaning of the vacuum pump to remove process deposits and replacement of the seals to prevent oil and vacuum leaks is essential to maintaining high performance of the product, as well as a clean and safe operating environment. A clean and overhaul kit contains all the parts needed for routine cleaning of your pump.

Blade (vane) replacement

Blades on a rotary pump will wear during normal operation. Edwards recommends replacing them every 30,000 hours (40 months), or as your process conditions dictate. A blade kit contains a full set of blades and typically is used in conjunction with a clean and overhaul kit.



OK - continue to monitor

Change oil immediately

Discontinue operation, overhaul pump and change oil

Accessories

If accessories are fitted to your pump, they too will require regular checks and maintenance. For example, a blocked dust filter can significantly reduce the efficiency of a pump.

Act responsibly and safely

Edwards cares about the health and safety of people and the environment. Remember to dispose of used or contaminated oil, cleaning solutions and components responsibly and safely.


Edwards services

If you prefer maintenance of your products to be performed with the assurances and guarantees of OEM standards, Edwards offers a full range of maintenance, overhaul and exchange services that include safe disposal of oil, fluids and used components. Contact us for further information.



service recommendations

Clean and overhaul kit: A37101131 (items in green below) Blade kit: A37101132 (items in red below)



Maintenance schedule interval	Change the oil every 3,000 hrs/4 months Mineral oil		Clean and overhaul every 15,000 hrs/20 months Clean & overhaul kit		Replace blades every 30,000 hrs/40 months Blade kit	
E2M0.7/1.5	H11026015		A37101131		A37101132	
Accessory maintenance	Check/char		nge filters	Check/	Check/change sorbents	
		Every month	Every 6 months	Refer to manual	Refer to manual	
		Odor/dust element	Oil element	Charcoal	Alumina	
Outlet mist filter	EMF 3	A22304081	A22304197	-	-	
Outlet mist filter	EMF 10	A22304079	A22304198	-	-	
Outlet mist filter	EMF 20	A22304077	A22304199	-	-	
Inlet chemical trap	ITC 20K	-	-	H12205001	H02600056	
Inlet dust filter	ITF 20K	A44201800	-	-	-	
Foreline trap	FL 20K	-	-	-	H02600056	

This is one of a range of datasheets provided by Edwards to help you check, maintain and optimise product performance. Please visit www.edwardsvacuum.com or contact us for other editions, accessory information or product specifications and test procedures.

Global contacts

EUROPE

UK Crawley +44 1293 528844 UK (local rate) 08459 212223 Belgium Brussels +32 2 300 0730 France Paris +33 1 4121 1256 Germany Munich 0800 000 1456 Italy Milan +39 02 48 4471



WWW.EDWARDSVACUUM.COM

USA Niagara (toll free) +1 800 848 9800 BRAZIL Sao Paulo +55 11 3952 5000 ISRAEL Qiryat-Gat +972 8 681 0633

ASIA PACIFIC

China (toll free) +86 400 111 9618 India, Pune +91 20 4075 2222 Japan, Yachiyo +81 47 458 8831 Korea, Bundang +82 31 716 7070 Singapore +65 6546 8408 Taiwan R.O.C. Jhunan Town +886 3758 1000

info@edwardsvacuum.com